


SPECIFICATION OF HIP PRESERVATION SET

LIST OF INSTRUMENTS IN HIP PRESERVATION SET

Department of Pediatric Orthopedics requires Hip Preservation Set, with technical specifications as follows -

Details of Instruments in the Set -

S.No.	Instrument with specification	Quantity
1	Straight Pelvic Osteotome 15 mm	1
2	Straight Pelvic Osteotome 20 mm	1
3	Curved Pelvic Osteotome 15 mm	1
4	Curved Pelvic Osteotome 20 mm	1
5	Double Curved Pelvic Osteotome 15 mm	1
6	Double Curved Pelvic Osteotome 20 mm	1
7	Flag Pelvic Osteotome 15 mm	1
8	Flag Pelvic Osteotome 20 mm	1
9	Pelvic Osteotome 15 mm	1
10	Pelvic Osteotome 20 mm	1
11	Curved Pelvic Osteotome 15 mm	1
12	Curved Pelvic Osteotome 20 mm	1
13	15 mm Straight Pelvic Osteotome	1
14	20 mm Short Curved Pelvic Osteotome	1
15	15 mm Pelvic Osteotome	1
16	20 mm Pelvic Osteotome	1
17	Hammer	1
18	3.5 mm Screw Driver	1
19	Malleable Retractor 300 mm long 25 mm width	2
20	Blunt Pelvic Retractor	2
21	Radiolucent Hohman Retractor 30 mm	6
22	Radiolucent Hohman Retractor 22 mm wide Tip	6
23	Spoon Retractor	1
24	Femoral Head Template 37 mm	1
25	Femoral Head Template 39 mm	1
26	Femoral Head Template 41 mm	1
27	Femoral Head Template 43 mm	1
28	Femoral Head Template 45 mm	1
29	Femoral Head Template 47 mm	1
30	Femoral Head Template 49 mm	1
31	Femoral Head Template 51 mm	1
32	Femoral Head Template 53 mm	1
33	2 mm x 9" guide wire	5
34	Osteotomy Angle measuring spreader Guide	2
35	Osteotomy Gap Measuring Guide	2
36	Osteotomy Spreader 220 mm, 10 mm width	2
37	Radio-opaque Triangles for measuring Femoral Osteotomy angles – 4 different sizes (40°/50°/90°) (30°/60°/90°) (20°/60°/100°) (30°/70°/80°)	1 each
38	Variable Angle Wire Guide for Femoral Osteotomy	1



39	Bending Iron for Recon Plates	1
40	Bending Pliers for Recon Plates	1 pair
41	In situ Bending Pliers for Recon Plates	1
42	In situ Bending Handle for Recon Plates	1 pair
43	Thin Long Osteotomy Chisels (with Gradation markings)– 10, 15, 20, 25 mm	2 each
44	DHS Guide Wire Indirect measuring device	1
45	One Large perforated Anodised aluminum / Stainless Steel Autoclaving Box with Modular Tray and silicone padding that can safely accommodate all the above instruments	2

Other Specifications

- 1- Up to 2 mm of the size difference between the specification of Retractors and Forceps from those being quoted is permissible.
- 2- The instruments should be of improved steel (Except malleable retractors) with high precision quality.

Conditions for tender:


1. All accessories should be from the same Original Equipment Manufacturer for the main unit.
2. Instruments must be ISO certified and a copy should be enclosed. (The ISO Certificate must be issued by any organization accredited by the Bureau of Indian Standard or accredited by the international accrediting forum "IAF" (Certificate to be attached).
3. Should be USA FDA and/ or European CE be approved by 4 digits notified body.
4. Other necessary certifications if any required will be provided by the bidder for the smooth functioning of the machine.
5. Installation process should be performed by O.E.M trained service engineers/ service representatives on OEM letterhead or Service Report within 15 days of supply, with the mandatory provision of providing preventive services visit of OEM trained Service Engineer/ Service Representative quarterly per year till the completion of warranty period (i.e., 20 visits for the first 05 years) & further quarterly visits (04 visits/year) year till the completion of CMC period.
6. The equipment should have a Brand name/ Model Number embossed/etched on the equipment.
7. All the technical specifications in the compliance statement must be supported by Original Literature from the firm/ O.E.M with highlighting Numbering & flagging of all technical certificates.
8. Offered Equipment should have a strong Government Installation base.

Prof. Avinash Agrawal
 MD, IJCC, IJCCM, FICCM, FIACCM, FCCP (USA)
 Head, Department of Critical Care
 King George's Hospital

9. Offered Equipment should have a Regional Sales Service Centre of the Original Equipment Manufacturer in the region for a 95 % uptime guarantee.
10. For the offered main unit, the essential, optional required consumables'/accessories' shelf life should be declared on the Original Equipment Manufacturer's letterhead.
11. In case of technical snag/failure/breakdown the response time for the inspection should be within 24 hours and repair within 05 days otherwise provide a service machine/alternate arrangement to be made till the period of recovery of the breakdown of the unit, failing which attracts penal action as per the decision of institute/ hospital.
12. For offered equipment the Training of technical staff and users should be performed by Original Equipment Manufacturer trained Service Engineers at the proper designated place- at bidders' cost.
13. Company should quote their latest model and need to provide an affidavit for the same.
14. As a tendering process the Demonstration of the offered Equipment is Mandatory at hospital/institute premises or other designated places at the bidder's cost.
15. The bidder must comply with the General Financial Rules and their modifications if any issued by the Government of India- 2017.
16. Any bidder from a country that shares a land border with India will be eligible to bid in the tender only if the bidder is registered with the Competent Authority (i.e., Registration certificate issued by the Ministry of Commerce and Industry (Department for Promotion of Industry and Internal Trade- DPIIT after October 2020). If any such bidder is not registered with DPIIT they will be liable for technical disqualification.
17. Principal (OEM) must authorize only one agent to be quoted in the bid otherwise multiple quotes through different agents in the same bid will be canceled.
18. The Bidder and its OEM both have to submit a notarized affidavit on the Indian Non-Judicial Stamp Paper of Rs.100/- that the bidder has not quoted the price higher than the current financial year and last financial year supplied to any government Institute/ Organization/ reputed Private Organization. OEM also has to submit that the price quoted by the bidder in the bid is on its behalf and the lowest in the current and last financial year in the country. Therefore, if at any stage it has been found that the supplier and its OEM have quoted lower rates than those quoted in this bid; the Institute (the purchaser) would be given the benefit of lower rates by the Supplier and any excess payment if any, will become immediately payable to the Institute (the purchaser). If such an affidavit is not submitted, the bid will be outrightly rejected. (Part of technical bid).

Prof. Avinash Agrawal
Head, Department of Critical care,
King George's Medical University, UP, Lko.

19. Guarantee / Warranty Period: Separate offers of Comprehensive Maintenance Contract (CMC on main equipment) and Annual Maintenance Contract (AMC on main equipment) for further 5 years after expiry of 5 years of warranty (i.e., 6th, 7th, 8th, 9th and 10th years) in rupees only (and on basis of percentage of price) should be included in a financial bid in the absence of which the offer is liable to be rejected. Payment for CMC/AMC shall be made only after the expiry of the warranty of 5 years, in case the Institute (the purchaser) decides for availing of CMC/AMC services. Contract for CMC/AMC shall be decided on expiry of warranty but rates (not more than 5% inclusive of all taxes for 6th to 10th year) will be frozen at the price of an issued purchase order before the release of payment by the Institute (the purchaser). However, the Institute (the purchaser) may decide not to enter into any CMC/AMC contract without assigning any reason for the same, which shall be binding upon the bid.



Prof. Avinash Agrawal
MD, IDCC, IFCCM, FICCM, FIACCM, FCCP (USA)
Head, Department of Critical care,
King George's Medical University, UP, Lko.

