SPECIFICATION FOR DRILL MACHINE

Handpiece

Speed around 60,000 rpm

Rotary direction switch on the handpiece

Safety off switch and cable position on the handpiece

Snap Lock for attachments with automatic locking

Machine washable

The weight of the handpiece should be around 185 gm

Power Should be around 80 Watts

Seal nipple made of metal should accommodate both ends of electric cable for cleaning

Electric Cable for Handpiece: should be preferably 4 m long

Hand Switch:

Removable

Continuously variable speed

Telescopic extension for better control of long attachments Lock slide for extra safety

Foot Switch (1 pedal):

Large pedal surface

Button for irrigation

Connection for additional foot switch

Bar for moving the foot switch

Continuously variable speed

Electric Cable for Foot Switch:

should be around 4 m

Option of Basic Console or Standard Console, with Irrigation:

Easy handling

Speed control at 4 levels

Integrated irrigation system with the setting of speed-dependent or constant

irrigation in standard Console

Irrigation pump with flow rate setting in standard Console

Connections for 2 drive units, I foot switch in standard console

Maintenance Kit:

Should have separate Adapters for Handpiece& Attachments

Should have high-pressure Maintenance Spray for internal cleaning of

July Jan la

Q-

Prof. Avinashage 1 of 5
MD, IDCC, IFCCM, FIGCM, FIACCM, FCCP (USA)
Head, Department of Critical care,
King George's Medical University, UP, Lko.

Hand piece & Attachments

Should have Synthetic bio-compatible oil for lubrication of moving parts

Drill Attachment with Mini Quick Coupling:

Speed around 1800 rpm approx

Drill Bits up to 1 mm to 6.0 mm in diameter

Drill Attachment with J-Latch coupling:

Speed around 1800 rpm approx

Drill Bits up to dia 6.0 mm

Oscillating Drill Attachment 45°, with Mini Quick Coupling:

Soft Tissue protection

Frequency: around 3200 OSC/min

Screw Attachment with Mini Quick Coupling:

Speed 400 rpm approx

Safe screw insertion with machine

Kirschner Wire Attachment:

Speed around 2700 rpm approx

Attachment for dia 0.6 1.6 mm Kirschner wires

Tensioning Lever can be rotated by 300°

Self-gripping

Drill Attachment 45°, cannulated, with Jacobs Chuck:

Speed 1800 rpm approx

Round drill bits can be used

Cannulation which is important e.g., inserting cannulated screws

Oscillating Saw Attachment:

Frequency: 22000 OSC/min approx

Amplitude: 15°

Sagittal Saw Attachment:

Frequency: 22000 OSC/min approx

Amplitude: 5.8"

Saw Blades can be inserted in 5 positions

Sagittal Saw Blade, Usable Length 18.0 mm to 31.0 mm, Blade width 2.5 mm to 6.0 mm,

Cutting Thickness 0.38 mm

Reciprocating Saw Attachment:

Frequency: 18000 OSC/min approx

la fr.

Prof. Rage 2nof 5rawal
MD, DCC JFCCM, FICCM, FIACCM, FCCP (USA)
Read, Department of Critical care,
King George's Medical University, UP, Lko.

A Comment of the Comm

Stroke: 2.5 mm

Craniotome Attachment and Dura Guards for Neurosurgery:

Gear ratio 1:1

Good turning capabilities

3 sizes of Dura Guards and corresponding Cranial Burrs

Option of Perforator with Hudson coupling or standard Perforator:

Gear reduction 97:1

For the safe opening of the cranial vault

Burrs – Should have sizes in the following ranges. Minor variations in the sizes and shapes are acceptable:

Burr, Round, dia 1.0 mm to 8.0 mm

Burr, Round, Diamond coated, dia 1.0 mm to 6.0 mm

Conditions for tenderer:

- 1. All accessories should be from the same Original Equipment Manufacturer for the main unit.
- Instruments must be ISO certified and a copy should be enclosed. (The ISO
 Certificate must be issued by any organization accredited by the Bureau of Indian
 Standard or accredited by the international accrediting forum "IAF" (Certificate to be
 attached).
- 3. Should be USA FDA and/ or European CE be approved by 4 digits notified body.
- 4. Other necessary certifications if any required will be provided by the bidder for the smooth functioning of the machine.
- 5. Installation process should be performed by O.E.M trained service engineers/ service representatives on OEM letterhead or Service Report within 15 days of supply, with the mandatory provision of providing preventive services visit of OEM trained Service Engineer/ Service Representative quarterly per year till the completion of warranty period (i.e., 20 visits for the first 05 years) & further quarterly visits (04 visits/year) year till the completion of CMC period.
- 6. The equipment should have a Brand name/ Model Number embossed/etched on the equipment.

See la R.

Prof. Avinash Agrawal
MD, IDCP, age: 3, of 5m, Deacom, Foor (USA)
Head Department of Critical care,
King George's Medical University, UP, I ko.

- 7. All the technical specifications in the compliance statement must be supported by Original Literature from the firm/ O.E.M with highlighting Numbering & flagging of all technical certificates.
- 8. Offered Equipment should have a strong Government Installation base.
- 9. Offered Equipment should have a Regional Sales Service Centre of the Original Equipment Manufacturer in the region for a 95 % uptime guarantee.
- 10. For the offered main unit, the essential, optional required consumables'/accessories' shelf life should be declared on the Original Equipment Manufacturer's letterhead.
- 11. In case of technical snag/failure/breakdown the response time for the inspection should be within 24 hours and repair within 05 days otherwise provide a service machine/ alternate arrangement to be made till the period of recovery of the breakdown of the unit, failing which attracts penal action as per the decision of institute/ hospital.
- 12. For offered equipment the Training of technical staff and users should be performed by Original Equipment Manufacturer trained Service Engineers at the proper designated place- at bidders' cost.
- 13. Company should quote their latest model and need to provide an affidavit for the same.
- 14. As a tendering process the Demonstration of the offered Equipment is Mandatory at hospital/institute premises or other designated places at the bidder's cost.
- 15. The bidder must comply with the General Financial Rules and their modifications if any issued by the Government of India- 2017.
- 16. Any bidder from a country that shares a land border with India will be eligible to bid in the tender only if the bidder is registered with the Competent Authority (i.e., Registration certificate issued by the Ministry of Commerce and Industry (Department for Promotion of Industry and Internal Trade- DPIIT after October 2020). If any such bidder is not registered with DPIIT they will be liable for technical disqualification.
- 17. Principal (OEM) must authorize only one agent to be quoted in the bid otherwise multiple quotes through different agents in the same bid will be canceled.
- 18. The Bidder and its OEM both have to submit a notarized affidavit on the Indian Non-Judicial Stamp Paper of Rs.100/- that the bidder has not quoted the price higher than

/an

Q.

Prof. Avipage 45075'al

MD, IDCC, IFCCM, FIGCM, FIGCM, FOCP (US.

Head, Department of Critical Care,

King George's Medical University (IP)

King thonge's interior On

2000

the current financial year and last financial year supplied to any government Institute/ Organization/ reputed Private Organization. OEM also has to submit that the price quoted by the bidder in the bid is on its behalf and the lowest in the current and last financial year in the country. Therefore, if at any stage it has been found that the supplier and its OEM have quoted lower rates than those quoted in this bid; the Institute (the purchaser) would be given the benefit of lower rates by the Supplier and any excess payment if any, will become immediately payable to the Institute (the purchaser). If such an affidavit is not submitted, the bid will be outrightly rejected. (Part of technical bid).

- 19. Guarantee / Warranty Period: Separate offers of Comprehensive Maintenance Contract (CMC on main equipment) and Annual Maintenance Contract (AMC on main equipment) for further 5 years after expiry of 5 years of warranty (i.e., 6th, 7th, 8th, 9th and 10th years) in rupees only (and on basis of percentage of price) should be included in a financial bid in the absence of which the offer is liable to be rejected. Payment for CMC/AMC shall be made only after the expiry of the warranty of 5 years, in case the Institute (the purchaser) decides for availing of CMC/AMC services. Contract for CMC/AMC shall be decided on expiry of warranty but rates (not more than 5% inclusive of all taxes for 6th to 10th year) will be frozen at the price of an issued purchase order before the release of payment by the Institute (the purchaser). However, the Institute (the purchaser) may decide not to enter into any CMC/AMC contract without assigning any reason for the same, which shall be binding upon the bid.
- 20. Should provide 5-year CMC. CMC cannot be more than 5% of the contract value.
- 21. System configured application-specific educational video tutorials shall be provided as standard with the system.
- 22. Details of service outlet in India to render services during 5 years warranty period.
- 23. The principals must give a certificate if the supplier/vendor is changed during the course of the guarantee/warranty period, the principals would be responsible for the upkeep/maintenance of the quote/supplied equipment, besides honouring all the terms and conditions of CMC/AMC in letter and spirit.

299

Jell

Prof. Avipage 5 of 5 val MD, IDCC, IFCCM, FICCM, FIACCM, FCCP (USA) Head, Department of Critical care, King George's Medical University, UP, Lko.