

Specifications of CRRT Machine

Operational Requirements	
S. No	
1.	The fully automated practice of a complete range of continuous renal replacement and fluid management therapies
2.	General features
2.1	S.C.U. F - Slow Continuous Ultra Filtration
2.2	C.V.V.H - Continuous Veno- Venous Hemofiltration
2.3	C.V.V.H. D - Continuous Veno-Venous Hemodialysis
2.4	C.V.V.H. D.F - Continuous Veno-Venous Hemodiafiltrations
2.5	T.P.E - Therapeutic Plasma Exchange
2.6	H.P. - Hemoperfusion
2.7	Sepsis treatment using special membranes (name the special membrane)
2.8	Compatibility with MARS monitor for Liver dialysis
3.	User-Friendly and has automated functions
3.1	Large Colour TFT-LCD touch screen (10 — 12 inches)
3.2	Smart software for easy operator guidance
3.3	Step-by-step instructions with graphical instructions on the screen for easy setup
3.4	Self-testing of alarms and functions after priming and every 1 to 2 hours to ensure the patient's safety
3.5	Rapid and automatic priming procedure (within 1- 7 minutes)
3.6	For continuous and precise fluid balance management dedicated (independent) weighing devices monitoring - at least four devices
	Weighing device - Pre-Blood pump

See

by

[Handwritten signature]

[Handwritten mark]

	Weighing device - Replacement pump
	Weighing device - Dialysate pump
	Weighing device - Effluent pump
	Each weighing scale should be able to accommodate up to 11 L of fluid at one time (a total of 44 for four scales)
3.7	Recording of patients' treatment history (up to 80 — 90 hours)
3.8	Events Storage (minimum of 450 — 500)
3.9	Recording and patient safety
	Total filtrate volume
	Replacement solution volume
	Dialysate volume
	Pre-blood solution volume
	Elapsed time
	Updated on the treatment history screen
3.10	Display of information on one screen
	Continuous information of all parameters
	Graphical display of pressure monitoring
3.11	Should operate with a low extracorporeal blood volume (equal or less than 152 ml, for pediatric 90-95 ml)
3.12	Built-in dosage calculator (tool to support operator on the dosage prescription and display dialysis dose delivered at end of therapy)
3.13	Option for easy changeover from one modality to another without interrupting the treatment
3.14	Option for Regional Citrate Anticoagulation for all therapies

See by


D. Johnson

13/7

3.15	Option for simultaneous delivery of Pre and Post filter replacement solution in CVVH and CVVHDF
3.16	Option for recirculation mode
3.17	Option to enable for Sepsis treatment with special compatible filter


Conditions for tender:

1. All accessories should be from the same Original Equipment Manufacturer for the main unit.
2. Instruments must be ISO certified and a copy should be enclosed. (The ISO Certificate must be issued by any organization accredited by the Bureau of Indian Standard or accredited by the international accrediting forum "IAF" (Certificate to be attached).
3. Should be USA FDA and European CE be approved by 4 digits notified body.
4. Other necessary certifications if any required will be provided by the bidder for the smooth functioning of the machine.
5. Installation process should be performed by O.E.M trained service engineers/ service representatives on OEM letterhead or Service Report within 15 days of supply, with the mandatory provision of providing preventive services visit of OEM trained Service Engineer/ Service Representative quarterly per year till the completion of warranty period (i.e., 20 visits for the first 05 years) & further quarterly visits (04 visits/year) year till the completion of CMC period.
6. The equipment should have a Brand name/ Model Number embossed/etched on the equipment.
7. All the technical specifications in the compliance statement must be supported by Original Literature from the firm/ O.E.M with highlighting Numbering & flagging of all technical certificates.
8. Offered Equipment should have a strong Government Installation base.
9. Offered Equipment should have a Regional Sales Service Centre of the Original Equipment Manufacturer in the region for a 95 % uptime guarantee.
10. For the offered main unit, the essential, optional required consumables'/accessories' shelf life should be declared on the Original Equipment Manufacturer's letterhead.
11. In case of technical snag/failure/breakdown the response time for the inspection should be within 24 hours and repair within 05 days otherwise provide a service machine/




alternate arrangement to be made till the period of recovery of the breakdown of the unit, failing which attracts penal action as per the decision of institute/ hospital.


12. For offered equipment the Training of technical staff and users should be performed by Original Equipment Manufacturer trained Service Engineers at the proper designated place- at bidders' cost.
13. Company should quote their latest model and need to provide an affidavit for the same.
14. As a tendering process the Demonstration of the offered Equipment is Mandatory at hospital/institute premises or other designated places at the bidder's cost.
15. The bidder must comply with the General Financial Rules and their modifications if any issued by the Government of India- 2017.
16. Any bidder from a country that shares a land border with India will be eligible to bid in the tender only if the bidder is registered with the Competent Authority (i.e., Registration certificate issued by the Ministry of Commerce and Industry (Department for Promotion of Industry and Internal Trade- DPIIT after October 2020). If any such bidder is not registered with DPIIT they will be liable for technical disqualification.
17. Principal (OEM) must authorize only one agent to be quoted in the bid otherwise multiple quotes through different agents in the same bid will be canceled.
18. The Bidder and its OEM both have to submit a notarized affidavit on the Indian Non-Judicial Stamp Paper of Rs.100/- that the bidder has not quoted the price higher than the current financial year and last financial year supplied to any government Institute/ Organization/ reputed Private Organization. OEM also has to submit that the price quoted by the bidder in the bid is on its behalf and the lowest in the current and last financial year in the country. Therefore, if at any stage it has been found that the supplier and its OEM have quoted lower rates than those quoted in this bid; the Institute (the purchaser) would be given the benefit of lower rates by the Supplier and any excess payment if any, will become immediately payable to the Institute (the purchaser). If such an affidavit is not submitted, the bid will be outrightly rejected. (Part of technical bid).
19. Guarantee / Warranty Period: Separate offers of Comprehensive Maintenance Contract (CMC on main equipment) and Annual Maintenance Contract (AMC on main equipment) for further 5 years after expiry of 5 years of warranty (i.e., 6th, 7th, 8th, 9th and 10th years) in rupees only (and on basis of percentage of price) should be included in a financial bid in the absence of which the offer is liable to be rejected. Payment for CMC/AMC shall be made only after the expiry of the warranty of 5 years, in case the







Institute (the purchaser) decides for availing of CMC/AMC services. Contract for CMC/AMC shall be decided on expiry of warranty but rates (not more than 5% inclusive of all taxes for 6th to 10th year) will be frozen at the price of an issued purchase order before the release of payment by the Institute (the purchaser). However, the Institute (the purchaser) may decide not to enter into any CMC/AMC contract without assigning any reason for the same, which shall be binding upon the bid.



Handwritten signatures and initials in blue ink, including a large signature, a signature with a long underline, and several smaller initials.